

File Revision Date:

October 10, 2017

Data Set Description:

PI: Richard Querel (For Station)
Glen McConville (for instrument)
Instrument: Dobson Ozone Spectrophotometer
Site(s): NIWA Lauder(45S, 170E)
Measurement Quantities: Total Column Ozone

Contact Information:

Name: Richard Querel
Address: Lauder, Private Bag 50061, Omakau, Central Otago, New Zealand
Phone: (+64) 3 440 0400
FAX: (+64) 21 072 2540
Email: Richard.Querel@niwa.co.nz

Contact for instrument information

Name: Glen McConville
Address: DOC/NOAA/OAR/ESRL R/GMD1, 325 Broadway, Boulder, Colorado, USA 80305
Phone: (+1) 303 497 3989
FAX: (+1) 303 497 5590
Email: Glen McConville

Reference Articles:

The instrument is described in numerous publications, the most commonly used reference is "Operations handbook - ozone observations with a Dobson spectrophotometer", W.D. Komhyr, Global Ozone Research and Monitoring Project. Report 183, World Meteorological Organization, Geneva, 2008.

Evans, R.D., Petropavlovskikh, I., McClure-Begley, A., McConville G., Quincy, D., and Miyagawa, K., The US Dobson Station network Data Record Prior to 2015, Re-evaluation of NDACC and WOUDC archived records with WinDobson Processing Software, Atmos. Chem. Phys., <https://doi.org/10.5194/acp-2017-383>, 2017.

Instrument Description:

Dobson Ozone Spectrophotometer number 72 (automated)

Algorithm Description:

Uses algorithm described in "Operations handbook - ozone observations with a Dobson spectrophotometer", W.D. Komhyr, Global Ozone Research and Monitoring Project. Report 183, World Meteorological Organization, Geneva, 2008.

www.esrl.noaa.gov/gmd/ozwv/dobson/GAW183-Dobson-WEB.pdf

Uses Bass/Paur ozone absorption coefficients, as defined in www.esrl.noaa.gov/gmd/ozwv/dobson/papers/coeffs.html

Expected Precision/Accuracy of Instrument:

There is a paper; "Review of the Dobson spectrophotometer and its accuracy", Reid E. Basher, Global Ozone Research and Monitoring Project. Report 13, World Meteorological Organization, Geneva, 1982, describing the precision and accuracy.

In general, the precision is considered to be from +/-1% (direct sun observations) to +/-5% (Observations on cloud zenith) for total ozone.

Accuracy is part of an ongoing debate, but is considered in the 5% range.

Instrument History:

1987.01.01-9999.99.99 ;D072