

File Revision Date:  
September 27, 2019

Data Set Description

PI: Monica Navarro Comas / Margarita Yela Gonzalez  
Instrument: UV-Visible Spectrometer RASAS  
Site: IZANA 28.308° N, 16.493° W  
Measurement Quantities: O3, NO2

Contact Information:

Name: Mónica Navarro Comas/Margarita Yela/Olga Puenteadura  
Address: Car. Ajalvir km4, 28850 Torrejon de Ardoz, Spain  
Phone: +34 91 5201258  
FAX: +34 91 5201317  
Email: navarrocm@inta.es/yelam@inta.es/puentero@inta.es

Reference Articles:

- Yela, M., Gil-Ojeda, M., Navarro-Comas, M., Gonzalez-Bartolomé, D., Puenteadura, O., Funke, B., Iglesias, J., Rodríguez, S., García, O., Ochoa, H., and Deferrari, G., Hemispheric asymmetry in stratospheric NO2 trends, *Atmos. Chem. Phys.*, 17, 13373-13389, <https://doi.org/10.5194/acp-17-13373-2017>, 2017.
- Robles-Gonzalez, C., Navarro-Comas, M., Puenteadura, O., Schneider, M., Hase, F., Garcia, O., Blumenstock, T., and Gil-Ojeda, M., Intercomparison of stratospheric nitrogen dioxide columns retrieved from ground-based DOAS and FTIR and satellite DOAS instruments over the subtropical Izana station, *Atmos. Meas. Tech.*, 9, 4471-4485, <https://doi.org/10.5194/amt-9-4471-2016>, 2016.
- Gil-Ojeda, M., Navarro-Comas, M., Gómez-Martín, L., Adame, J. A., Saiz-Lopez, A., Cuevas, C. A., González, Y., Puenteadura, O., Cuevas, E., Lamarque, J.-F., Kinninson, D., and Tilmes, S., NO2 seasonal evolution in the north subtropical free troposphere, *Atmos. Chem. Phys.*, 15, 10567-10579, doi:10.5194/acp-15-10567-2015, 2015.
- Gomez, L., Navarro-Comas, M., Puenteadura, O., Gonzalez, Y., Cuevas, E., and Gil-Ojeda, M., Long-path averaged mixing ratios of O3 and NO2 in the free troposphere from mountain MAX-DOAS, *Atmos. Meas. Tech.*, 7, 3373-3386, <https://doi.org/10.5194/amt-7-3373-2014>, 2014.
- Puenteadura, O., Gil, M., Saiz-Lopez, A., Hay, T., Navarro-Comas, M., Gómez-Pelaez, A., Cuevas, E., Iglesias, J., and Gomez, L., Iodine monoxide in the north subtropical free troposphere, *Atmos. Chem. Phys.*, 12, 4909-4921, doi:10.5194/acp-12-4909-2012, 2012.
- Roscoe, H.K., et al., Intercomparison of slant column measurements of NO2 and O4 by MAX-DOAS and zenith-sky UV and visible spectrometers, *Atmos. Meas. Tech.*, 3, 1629-1646, 2010.
- Gil, M., Yela, M., Gunn, L. N., Richter, A., Alonso, I., Chipperfield, M. P., Cuevas, E., Iglesias, J., Navarro, M., Puenteadura, O., and Rodríguez, S.: NO2 climatology in the northern subtropical region: diurnal, seasonal and interannual variability, *Atmos. Chem. Phys.*, 8, 1635-1648, doi:10.5194/acp-8-1635-2008, 2008.

Instrument description:

Name: RASAS\_02

Location: Indoor

Spectrometer type: Shamrok SR-163i spectrograph

Grating: Holographic 1200 grooves/mm blazed a 300nm

Detector: 1024x255 pixels DU420A-BU Andor Idus CCD

Input optic: 10m Quartz fiber optic pointing at IEA: 90°, 70°, 30°, 10°, 5°, 3°, 2°, 1°, 0°, -1°

Detector Temperature: -30°C (Peltier + circulating cooler)

Housing temperature: +18° ± 0.2° C

Wavelength region: 410-525 nm

Field of view: 1°

Spectral resolution: 0.6 nm

Sampling ratio: 5 samples/FWHM

Linear dispersion: 0.11 nm/pixel

Instrument automatic control: Home made

Name: RASAS\_01

Location: Indoor

Spectrometer type: Jarrell Ash Monospec 18

Grating: 600 g/mm ruled

Input optic: 5m Quartz fiber optic pointing the sky

Detector: EG&G 1453<sup>a</sup> with a 1024 Reticon PDA

Detector controller: EG&G 1461

Detector Temperature: -40°C (Peltier + circulating cooler)

Housing temperature: +18° ± 0.2° C

Wavelength region: 350-590 nm

Field of view: 10°

Spectral resolution: 1.2 nm

Pixel size: 0.257 nm

Sampling ratio: 5 samples/FWHM

ADC: 14 bits

Instrument automatic control: Home made

Name: EVA\_01

Spectrometer type: Jobin-Yvon H20

Location: Outdoor

Grating: 1200 g/mm holographic

Input optic: Mirror 45°

Detector: PMT in current mode

Detector controller: Home made

Detector Temperature: Room

Housing temperature: Room

Wavelength region: 430-450 nm

Field of view: 10°

Spectral resolution: 1 nm

Samples: 10/nm

Sampling ratio: 10 samples/FWHM  
ADC: 16 bits  
Instrument automatic control: Home made

Algorithm description:

Optical depths calculated as the log of the ratio of a reference high sun spectrum with the measured spectrum are fitted to laboratory cross-sections using a least square method. Stretching and shifting are taken into account for the fit.

Cross-sections of NO<sub>2</sub>, O<sub>3</sub>, O<sub>4</sub>, H<sub>2</sub>O, and Rayleigh curvature are included in the analysis. Ring is corrected by including a pseudo-cross section in the fitting process.

The amount in the reference spectra are estimated by Langley plots (O<sub>3</sub>) and iterative approximation using twilight am and pm (NO<sub>2</sub>)

Dark current is calculated from the integration time accounting by interpixel variability. DC measurements are carried out under routine basis by an electronic shutter located close to the optics. Spectral ranges used for standard analysis are: RASAS\_01: 450-530 nm for NO<sub>2</sub> and O<sub>3</sub>; RASAS\_02: 430-520 nm for NO<sub>2</sub> and O<sub>3</sub>; EVA: 430-450 nm for NO<sub>2</sub>

Expected precision/Accuracy of Instrument:

The error budget to each measured value, discriminate random and systematic error sources.

Random error is dominated by the uncertainties related to the slant column spectral fit (due to detector noise, instrumental imperfections, as well as errors or unknowns in the signal modeling) and the calculations of the Air Mass factors (errors related to the choice of the radiative transfer model settings, i.e. the O<sub>3</sub> and NO<sub>2</sub> vertical profiles, the aerosol extinction profile, the cloud conditions, and in case of NO<sub>2</sub>, the inclusion or not of the rapid twilight photochemistry).

The systematic error budget is dominated by the uncertainties of the O<sub>3</sub> and NO<sub>2</sub> cross sections used in the spectral fit and the uncertainty on the determination of the residual amount of O<sub>3</sub> and NO<sub>2</sub> in the reference spectra by using the Langley-plot technique.

The estimated overall errors in the individual measurements are, on average, approximately:

for NO<sub>2</sub>: 1 % fit analysis; 5 % AMF; 2 % cross-sections; 2 % residual column.

for O<sub>3</sub>: 0.5 % fit analysis; 3.6 % AMF; 3 % cross-sections; 2 % residual column.

Total random and systematic uncertainty for each measured value are given in the data.

Instrument History:

EVA and RASAS\_01 instruments have been operating together since 1999 to 2010 and EVA and RASAS\_02 since 2010 for overlapping purposes.

Data submitted to the database are:

2010-present	RASAS_02: NO <sub>2</sub> O <sub>3</sub>
1999-2010	RASAS_01: NO <sub>2</sub> O <sub>3</sub>
1993-1999	EVA_0: NO <sub>2</sub>

RASAS\_02 had been tested in the Cabauw Intercomparison of Nitrogen Dioxide measuring Instruments (CINDI) campaign in 2009 in Cabauw, the Netherlands.

RASAS\_01 meet the certification criterium for type 2 instruments in the blind NDSC intercomparison of 1996 (OHP).

EVA NO2 has been compared with RASAS\_01/RASAS\_02. The agreement between instruments is within 5% (1 sigma).